



Accuracy report for Jamaica Bottle Filling Plant with 18,000 gallon tank, in metric

Note: operations stop at 4PM (16:00) and Start again at 08:00 (am)

Net Truck scale weight of product delivered	19560	Note: Scale is used as constant, with accepted error potential	
Tank total METRIC (KG) capacity	39054		

	LEVEL	PRESSURE	NET FLUID VOL	NET VAPOR VOL	TOTAL NET VOL	TOTAL MASS	DATE	TIME	Gauge KG of delivery	Delivery Delta to scale	% deviation on delivery	Deviation on full scale (FTI advertises 1/2 of 1%)
Last reading before delivery began	45.200	399	6,791	704	7,495	4,294	10:11:2005	18:47				
First reading after delivery completed	161.500	372	41,192	276	41,468	23,761	10:11:2005	21:13	19,467	-93	-0.4755%	-0.2381%
Average of last 4 readings prior to morning operations (maximum settling time)	161.500	358	41,260	276	41,536	23,800	10:12:2005	07:26	19517	-43	0.2198%	-0.1107%
	161.600	358	41,276	272	41,548	23,806	10:12:2005	07:12				
	161.700	358	41,272	280	41,552	23,809	10:12:2005	07:39				
Last reading before operations, greatest settling time	161.900	358	41,298	287	41,585	23,828	10:12:2005	07:52	19534	-26	-0.1329%	-0.0666%